


America

# CERTIFICATE

No. U8 095184 0043 Rev. 00

**Holder of Certificate:** **E. ZOLLER GmbH & Co. KG**  
 Gottlieb-Daimler-Str. 19  
 74385 Pleidelsheim  
 GERMANY

**Certification Mark:**



**Product:** **Laboratory equipment**

This product was voluntarily tested to the relevant safety requirements referenced on this certificate. It can be marked with the certification mark above. The mark must not be altered in any way. This product certification system operated by TÜV SÜD America Inc. most closely resembles system 3 as defined in ISO/IEC 17067. Certification is based on the TÜV SÜD "Testing and Certification Regulations". TÜV SÜD America Inc. is an OSHA recognized NRTL for USA and a Standards Council of Canada ISO/IEC 17065 accredited Certification body for Canada.

**Test report no.:** 713235451

**Date,** 2022-11-04

*H. Elze*  
 ( Hendrik Elze )



America

# CERTIFICATE

No. U8 095184 0043 Rev. 00

**Model(s):** powerShrink 600

**Tested according to:** UL 61010-1:2012/R:2019-07  
CSA C22.2 No. 61010-1:2012/A1:2018-11

## Parameter:

Rated voltage: 480Y/277 V (+10, -4 %), (3Ph + PE)  
Rated frequency: 60 Hz  
Rated power: 15 kVA  
Protection class: 1

Ambient temp.: 5 to 30 °C  
Air pressure: 6-8 bar  
Pollution degree: 2  
Overvoltage category: II

## Conditions of acceptability:

- This equipment is for indoor use in non-hazardous locations, operated by qualified personnel skilled in its use.
- The detachable power cord set shall be complying with the National Standards and/or National Regulations of the country in question and/or in compliance with this standard. The detachable power cord set shall not longer than 3 m.