


America

CERTIFICATE

No. U8 095184 0036 Rev. 01

Holder of Certificate: **E. ZOLLER GmbH & Co. KG**
 Gottlieb-Daimler-Str. 19
 74385 Pleidelsheim
 GERMANY

Certification Mark:



Product: **Electronic measuring equipment**

This product was voluntarily tested to the relevant safety requirements referenced on this certificate. It can be marked with the certification mark above. The mark must not be altered in any way. This product certification system operated by TÜV SÜD America Inc. most closely resembles system 3 as defined in ISO/IEC 17067. Certification is based on the TÜV SÜD "Testing and Certification Regulations". TÜV SÜD America Inc. is an OSHA recognized NRTL for USA and a Standards Council of Canada ISO/IEC 17065 accredited Certification body for Canada.

Test report no.: 713274919

Date, 2022-12-06

(Ralph Fischer)



America

CERTIFICATE

No. U8 095184 0036 Rev. 01

Model(s): toolBalancer H550, toolBalancer H750

Brand Name: ZOLLER

Tested according to: UL 61010-1:2012/R2019-07
 CSA C22.2 No. 61010-1:2012/A1:2018-11
 CSA E60825-1:2015

Parameters:

| | |
|-----------------------|---|
| Rated voltage: | 120-230 V |
| Rated frequency: | 50/60 Hz |
| Rated power: | 720 VA |
| Protection class: | I |
| Overvoltage category: | II |
| Pollution degree: | 2 |
| Air pressure: | 6-8 bar |
| Ambient: | restricted: 5 to 35 °C, 10 to 90 % r.H. (non-condensing) |
| Laser: | |
| Wavelength: | 650 nm |
| Laser class: | 1 |
| Output power: | 5 mW |