


America

CERTIFICATE

No. U8 095184 0045 Rev. 00

Holder of Certificate: **E. ZOLLER GmbH & Co. KG**
 Gottlieb-Daimler-Str. 19
 74385 Pleidelsheim
 GERMANY

Certification Mark:




Product: **Laboratory equipment**

This product was voluntarily tested to the relevant safety requirements referenced on this certificate. It can be marked with the certification mark above. The mark must not be altered in any way. This product certification system operated by TÜV SÜD America Inc. most closely resembles system 3 as defined in ISO/IEC 17067. Certification is based on the TÜV SÜD "Testing and Certification Regulations". TÜV SÜD America Inc. is an OSHA recognized NRTL for USA and a Standards Council of Canada ISO/IEC 17065 accredited Certification body for Canada.

Test report no.: 713294501

Date, 2023-05-17

(Erich Thurner)



America

CERTIFICATE

No. U8 095184 0045 Rev. 00

Model(s): »powerShrink 400« zStream, »powerShrink 400«

Tested according to: UL 61010-1:2012/R:2019-07
CSA C22.2 No. 61010-1:2012/A1:2018-11

Parameters:

Rated voltage:	480Y/277 V (-10, +4 %), (3Ph + PE)
Rated frequency:	60 Hz
Rated power:	15 kVA
Protection class:	I
Ambient temp.:	5 to 30 °C
Air pressure:	6-8 bar
Pollution degree:	2
Overvoltage category:	II

Remarks/CoA (e.g. nomenclature; conditions of acceptability):

- This equipment is for indoor use in non-hazardous locations, operated by qualified personnel skilled in its use.
- The mains voltage fluctuation is -10 % and +4 % only. During the installation the mains voltage fluctuation shall be considered on site.